
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ4P00AZ.02
---	---	-------------

Fletcher Trait Plasma, 2x1mL

REF 5205006
LOT 4P31B01.01
 28-Feb-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.9	7.6	OK

Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B00	78.0 – 130.0%	106.5	OK
	CCA LOT 3P38B00	34.5 – 57.5%	42.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Monika WEINBAUER

27-Mar-2015

Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

27-Mar-2015

Datum / Date: