
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1R00AZ
---	---	----------

Coagulation Reference, 5x1mL

REF 5220110
LOT 1R31B00.02
 31-Oct-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	DAPTTIN TC, sec	30.0 – 40.6	37.1	OK
	FVIII Activity, (%)	63.0 – 105.0	90.4	OK

**Durchführung von
Prepared by**



30. Jul. 2013
Datum / Date:

**Freigabe von
Approved by**

Harald SEVCIK


30. JULI 2013
Datum / Date: