



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1R00AZ.01


Coagulation Reference 5x1mL

REF 5220110
LOT 1R2EB00.06
⌚ 30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Verwendbarkeit Usability Ceveron® alpha	CCN (1P14B00) Factor X	81 – 135 %	114.6 %	OK
	CCA (3P11B00) Factor X	14 – 24 %	23.6 %	OK

Durchführung von
Prepared by

Alessandro BENIGNI



15-Apr-2013
Datum / Date:

Freigabe von
Approved by

Harald SEVCIK



16. APR. 2013
Datum / Date: