
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ1R00AZ
---	---	----------

**Coagulation Reference, 5x1mL**

REF      5220110  
LOT      1R31B00.02  
      31-Oct-2016

<b>Analytische Prozedur Analytical Procedure</b>		<b>Anforderung Requirement</b>	<b>Ergebnis Result</b>	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron <sup>®</sup> alpha	DAPTTIN TC, sec	30.0 – 40.6	37.1	OK
	FVIII Activity, (%)	63.0 – 105.0	90.4	OK

**Durchführung von  
Prepared by**

  
 \_\_\_\_\_

30. Jul. 2013  
 Datum / Date:

**Freigabe von  
Approved by**

**Harald SEVCIK**  
  
 \_\_\_\_\_

30. JULI 2013  
 Datum / Date: