

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ1R00AZ
---	---	----------

### Coagulation Reference, 5x1mL

REF      5220110  
LOT      1R31B00.03  
      31-Oct-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron <sup>®</sup> alpha	Technoclot <sup>®</sup> PT Plus, (%)	94.3 – 127.8	107.7	OK
	Technoclot <sup>®</sup> PT Plus, (INR)	0.82 – 1.12	0.98	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


Durchführung von  
Prepared by

**Lavdrim NASUFI**  
  


---

26-Feb-2014  
Datum / Date:

Freigabe von  
Approved by

**Lucas HOLOVICZ**  
  


---

26-Feb-2014  
Datum / Date: