



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1R00AZ

Coagulation Reference, 5x1mL

REF 5220110

LOT 1R51B00.02

⌚ 31-Oct-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1.2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen, Clauss meth. CCN LOT 1P37B00	2.52 – 3.78 g/L	3.44	OK
	Fibrinogen, Clauss meth. CCA LOT 3P38B00	0.93 – 1.39 g/L	1.11	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Lucas ROSENMAYR

Datum / Date:

Freigabe von  
Approved by

Monika WEINBAUER

Datum / Date: