

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1R00AZ
---	---	----------

Coagulation Reference, 50x1mL

REF 5220120
LOT 1R31B00.01
 31-Oct-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	DAPTTIN TC, sec	30.0 – 40.6	37.1	OK
	FVIII Activity, (%)	63.0 – 105.0	90.4	OK

Durchführung von **Lavdrim NASUFI**
Prepared by



20. Juni 2013
Datum / Date:

Freigabe von **Harald SEVCIK**
Approved by



20. JUNI 2013
Datum / Date: