



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1R00AZ

Coagulation Reference, 50x1mL

REF 5220120

LOT 1R51B00.01

⌚ 31-Oct-2016

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|---|----------------------------------|----------------------------------|----|
| pH | | 7.0 – 7.8 | 7.6 | OK |
| Virologie Virology | HbsAg HIV 1.2 Ab HCV | Negative Negative Negative | Negative Negative Negative | OK |
| Verwendbarkeit Usability Ceveron® alpha | Fibrinogen, Clauss meth. CCN [LOT] 1P37B00 | 2.52 – 3.78 g/L | 3.31 | OK |
| | Fibrinogen, Clauss meth. CCA [LOT] 3P38B00 | 0.93 – 1.39 g/L | 1.19 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

21-May-2015
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI *Lavdrim Nasufi*

21-May-2015
Datum / Date: