
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1R00AZ
---	---	----------

Ceveron® Coagulation Reference, 5x1mL

REF 5220130
LOT 1R51B00.01
 31-Oct-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen, Clauss meth. CCN LOT 1P37B00	2.52 – 3.78 g/L	3.68	OK
	Fibrinogen, Clauss meth. CCA LOT 3P38B00	0.93 – 1.39 g/L	1.17	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

21-May-2015
Datum / Date:

**Freigabe von
Approved by**

Lavdrin Nasufi
Lavdrin NASUFI

21-May-2015
Datum / Date: