
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9C00AZ
---	---	----------

Ca-Chloridlösung 25mmol/L, 25mL

REF 5277017
LOT 9C31BCV.02
 31-Dez-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Sterilität sterility		0 CFU/mL	0	OK	
CaCl ₂ -Gehalt CaCl ₂ -Content		23.5 – 26.5 mmol/L	24.7	OK	
Verwendbarkeit Usability Ceveron [®] alpha	Daptin	TCN LOT 1A31B01	32.7 – 40.3 sec	37.1	OK
		TCA LOT 3P31B01	87.4 – 107.3 sec	102.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by

Alessandro BENIGNI

21-Feb-2014

Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

21-Feb-2014

Datum / Date: