

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9C00AZ.02
---	---	-------------

Ca-Chloridlösung 25mmol/L 25mL

REF 5277017
LOT 9C53BB0.01
 31-Mar-2020


Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result			
CaCl ₂ Gehalt CaCl ₂ Content	23.5 – 26.5 mmol/L	26.1	OK		
Verwendbarkeit, Usability Ceveron [®] alpha	Siron LS	CCN LOT 1P37B00	24.9 – 30.7 sec	28.3	OK
		TCA LOT 3P38B00	62.3 – 76.9 sec	69.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by


Lucas ROSENMAYR

30-Apr-2015
Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

30-Apr-2015
Datum / Date: