

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ8B00AZ
---	---	----------

**Technochrom AT III Sub. Th-1 2x10µmol**

REF      5340011  
LOT      8B41B00.03  
      31-May-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Chromogenes Substrat: F IIa Chromogenic substrate: F IIa	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P44B00	72.0 – 120.0%	109.7    OK
	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	28.5 – 47.5%	39.8    OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

**Durchführung von  
Prepared by**

  
**Lavdrim NASUFI**

25-Feb-2015  
Datum / Date:

**Freigabe von  
Approved by**

  
**Alessandro BENIGNI**

25-Feb-2015  
Datum / Date: