
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ8B00AZ.01
---	---	-------------

**Technochrom AT III Sub. Th-1 2x10µmol**

REF      5340011  
LOT      8B51B00.01  
      31-Mar-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability Chromogenes Substrat: F IIa Chromogenic substrate: F IIa	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P37B00	72.8 – 121.3%      104.1      OK
	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	28.5 – 47.5%      40.5      OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK

**Durchführung von  
Prepared by**

  
Lavdun NASUFI

08. MRZ. 2016  
\_\_\_\_\_  
Datum / Date:

**Freigabe von  
Approved by**

  
Christoph ENNE

08-Mar-2016  
\_\_\_\_\_  
Datum / Date: