
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7S00AZ.01
---	---	-------------

### Technochrom AT III Reagent A2 2x43 IE

REF      5340217  
LOT      7S11B01.02  
      31-Jan-2016


Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	8.2 – 8.6	8.4	OK
Verwendbarkeit Usability Ceveron® alpha	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P44B00	72.0 – 120.0%	97.6    OK
	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	28.5 – 47.5%	42.4    OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

**Durchführung von**  
**Prepared by**

Alessandro BENIGNI  


09-Feb-2015  
Datum / Date:

**Freigabe von**  
**Approved by**

Lavdrim NASUFI  


09-Feb-2015  
Datum / Date: