




| | | |
|---|---|-------------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ0800AZ.01 |
|---|---|-------------|


TECHNOCHROM® anti-Xa Kit


REF 5340250
LOT 0841B00.01
 31-Mar-2016

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|--------------------|---|---|
| 1x25mL | R1 anti Xa Buffer | 9R44B00 | 28-Feb-2019 |
| 1x4mL | R2 FXa | 7K44B00 | 31-Mar-2016 |
| 1x4mL | R3 Substrat | 8A44B00 | 31-Mar-2016 |


| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|--|----------------------------|--------------------|----|
| Verwendbarkeit Usability Ceveron® alpha | Apixaban Control Low LOT 3932B00 | 98.9 – 164.8 ng/mL | 151.6 | OK |
| | Apixaban Control High LOT 3D32B00 | 253.5 – 422.5 ng/mL | 379.9 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |


Durchführung von
Prepared by

Lucas HOLOVICZ



Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI



Datum / Date: