

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0N00.01
---	---	-----------


Technochrom® F VIII:C Reagent Kit, 40 T.

REF 5344101
LOT 0N41B00.01
 30-Apr-2015

Inhalt Content	Reagenz Reagent	LOT	
2x2mL	Substrate for factor Xa 10µmol	8M41B00	31-Dec-2018
2x2mL	F VIII:C Reagent A	7M04B02	30-Nov-2015
2x2mL	F VIII:C Reagent B	7L33B00	31-May-2015
1x1mL	Reference Standard 1, F VIII:C	2W21B01	30-Nov-2015
1x1mL	Reference Standard 2, F VIII:C	2X21B01	30-Nov-2015
1x1mL	Reference Standard 3, F VIII:C	2Y21B01	30-Nov-2015
1x1mL	Reference Standard 4, F VIII:C	2Z31B0D	30-Apr-2015
2x8mL	FVIII:C Reaction buffer	9T31B00	31-Oct-2018
3x30mL	FVIII:C Dilution buffer	9Y31B00	31-Oct-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P14B00	72 - 120%	104.6 OK
	CCA LOT 3P24B00	14.3 – 23.8%	19.3 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Lavdrim NASUFI


03-Mär-2014
Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI


03-Mär-2014
Datum / Date: