

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0N00.02
---	---	-----------

**TECHNOCHROM® F VIII:C 40T.**

REF      5344101  
LOT      0N51B00.01  
      31-May-2016


Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
2x2mL	Substrate for factor Xa 10µmol	8M41B00	31-Dec-2018
2x2mL	F VIII:C Reagent A	7M04B03	30-Apr-2017
2x2mL	F VIII:C Reagent B	7L41B00	31-Aug-2016
1x1mL	Reference Standard 1, F VIII:C	2W21B02	30-Apr-2017
1x1mL	Reference Standard 2, F VIII:C	2X21B02	30-Apr-2017
1x1mL	Reference Standard 3, F VIII:C	2Y41B00	31-May-2016
1x1mL	Reference Standard 4, F VIII:C	2Z31B0E	31-Aug-2016
2x8mL	FVIII:C Reaction buffer	9T31B00	31-Oct-2018
3x30mL	FVIII:C Dilution buffer	9Y31B00	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN <span style="border: 1px solid black; padding: 2px;">LOT</span> 1P37B00	69.0 – 115.0%	89.2	OK
	CCA <span style="border: 1px solid black; padding: 2px;">LOT</span> 3P38B00	19.5 – 32.5%	27.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


**Durchführung von  
Prepared by**

**Fikret NASUFI**  
  


---

  
Datum / Date:

**Freigabe von  
Approved by**

  
**Lavdrim NASUFI**  


---

  
Datum / Date: