

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0J00.01
---	---	-----------

TECHNOCHROM® C1-INH REAGENT KIT

REF 5345003
LOT 0J37B00.02
 31-Jul-2015

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N f. C1-INH	1P14B00	30-Jun-2016
1x1mL	Coagulation Control A f. C1-INH	3P24B00	30-Jun-2016
1x1mL	Coagulation Reference f. C1-INH	1R31B00	31-Oct-2016
1x3mL	C1 Esterase	7B33B00	31-Jul-2015
1x3mL	Substrate C1-INH	8C32B00	31-Jul-2018
1x25mL	Sample Buffer A	9M31B00	31-Jul-2018
1x20mL	Reaction Buffer B	9N31B00	31-Jul-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN (1P14B00)	73-122%	117.4%	OK
	CCA (3P24B00)	29.8-49.6%	43.3%	OK

Durchführung von
Prepared by

Lucas HOŁOVIĆZ

21-AUG-2013
Datum / Date:

Freigabe von
Approved by

Harald SEVČIK

21. AUG. 2013
Datum / Date: