



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0J00.01

TECHNOCHROM® C1-INH REAGENT KIT

REF 5345003

LOT 0J38B00.01

🕒 30-Sep-2015

Inhalt Content	Reagenz Reagent	LOT	🕒
1x1mL	Coagulation Control N f. C1-INH	1P14B00	30-Jun-2016
1x1mL	Coagulation Control A f. C1-INH	3P24B00	30-Jun-2016
1x0,896mL	Coagulation Reference f. C1-INH	1R31B00	31-Oct-2016
1x3mL	C1 Esterase	7B37B00	30-Sep-2015
1x3mL	Substrate C1-INH	8C32B00	31-Jul-2018
1x25mL	Sample Buffer A	9M31B00	31-Jul-2018
1x20mL	Reaction Buffer B	9N31B00	31-Jul-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability Ceveron® alpha	CCN (1P14B00)	73-122% 119.1% OK
	CCA (3P24B00)	29.8-49.6% 46.9% OK

Durchführung von
Prepared by

Lucas HOLOVICZ

23-Oct-2013
Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI

23-Oct-2013
Datum / Date: