

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0J00AZ.01
-----------------------------------------------------------------------------------	-------------------------------------------------------	-------------

TECHNOCHROM® C1-INH REAGENT KIT

REF 5345003
LOT 0J41B00.03
 31-Dec-2015

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N f. C1-INH	1P44B00	31-May-2016
1x1mL	Coagulation Control A f. C1-INH	3P24B00	30-Jun-2016
1x0,825mL	Coagulation Reference f. C1-INH	1R31B00	31-Oct-2016
1x3mL	C1 Esterase	7B38B00	31-Dec-2015
1x3mL	Substrate C1-INH	8C41B00	30-Nov-2018
1x25mL	Sample Buffer A	9M31B00	31-Jul-2018
1x20mL	Reaction Buffer B	9N31B00	31-Jul-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P44B00	79.5 - 132.5%	106.6%	OK
	CCA LOT 3P24B00	28.5 - 47.5%	45.6%	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Alessandro BENICINI

16-Apr-2014

Datum / Date:

**Freigabe von
Approved by**

Lucas HOLOVICZ

16-Apr-2014

Datum / Date: