

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0J00AZ.01
---	---	-------------

TECHNOCHROM® C1-INH REAGENT KIT

REF 5345003
LOT 0J43B00.02
 31-Dec-2015

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N f. C1-INH	1P44B00	31-May-2016
1x1mL	Coagulation Control A f. C1-INH	3P38B00	30-Sep-2017
1x0,825mL	Coagulation Reference f. C1-INH	1R31B00	31-Oct-2016
1x3mL	C1 Esterase	7B38B00	31-Dec-2015
1x3mL	Substrate C1-INH	8C42B00	31-May-2019
1x25mL	Sample Buffer A	9M41B00	31-Mar-2019
1x20mL	Reaction Buffer B	9N41B00	31-Mar-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P44B00	79.5 - 132.5%	97.3	OK
	CCA LOT 3P38B00	29.3 – 48.8%	48.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

10-Sep-2014
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI

10-Sep-2014
Datum / Date: