





**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQ0J00AZ.01

TECHNOCHROM® C1-INH REAGENT KIT

REF 5345003
LOT 0J44B00.02
 31-May-2016

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N f. C1-INH	1P44B00	31-May-2016
1x1mL	Coagulation Control A f. C1-INH	3P38B00	30-Sep-2017
1x0,825mL	Coagulation Reference f. C1-INH	1R31B00	31-Oct-2016
1x3mL	C1 Esterase	7B41B00	30-Jun-2016
1x3mL	Substrate C1-INH	8C42B00	31-May-2019
1x25mL	Sample Buffer A	9M41B00	31-Mar-2019
1x20mL	Reaction Buffer B	9N41B00	31-Mar-2019


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P44B00	79.5 - 132.5%	104.8	OK
	CCA LOT 3P38B00	29.3 – 48.8%	39.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lavdrim NASUFI

03-Dec-2014
Datum / Date:

**Freigabe von
Approved by**


LUCAS HOLOVICZ

03-Dec-2014
Datum / Date: