

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0J00AZ.01
---	---	-------------

TECHNOCHROM® C1-Inh. KIT 30T.

REF 5345003
LOT 0J52B00.01
 31-Oct-2016

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N	1P37B00	30-Sep-2017
1x1mL	Coagulation Control A	3P38B00	30-Sep-2017
1x0,825mL	Coagulation Reference	1R51B00	31-Oct-2016
1x3mL	C1 Esterase	7B51B00	30-Apr-2017
1x3mL	Substrate f. C1-esterase	8C51B00	31-Mar-2020
1x25mL	C1-Inh Probenpuffer A	9M41B00	31-Mar-2019
1x20mL	C1-Inh Reaktionspuffer B	9N41B00	31-Mar-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B00	77.3 - 128.8%	96.2	OK
	CCA LOT 3P38B00	29.3 – 48.8%	37.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

01-Jun-2015
Datum / Date:

Freigabe von
Approved by

Laverim Nasufi
Laverim NASUFI

01-Jun-2015
Datum / Date: