



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ0J00AZ.01

TECHNOCHROM® C1-Inh. KIT 30T.

REF 5345003
LOT 0J53B00.02
 31-Oct-2016

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N	1P37B00	30-Sep-2017
1x1mL	Coagulation Control A	3P38B00	30-Sep-2017
1x0,825mL	Coagulation Reference	1R51B00	31-Oct-2016
1x3mL	C1 Esterase	7B51B00	30-Apr-2017
1x3mL	Substrate f. C1-esterase	8C51B00	31-Mar-2020
1x25mL	C1-Inh Probenpuffer A	9M51B00	30-Jun-2020
1x20mL	C1-Inh Reaktionspuffer B	9N51B00	31-Jul-2020

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability	CCN LOT 1P37B00	77.3 – 128.8%	85.9	OK
Ceveron® alpha	CCA LOT 3P38B00	29.3 – 48.8%	37.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von
Prepared by

Christoph Enne
Christoph ENNE

12. NOV. 2015

Datum / Date:

Freigabe von
Approved by

Alessandro Benigni
Alessandro BENIGNI

13-Nov-2015
Datum / Date: