



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0J00AZ.01

TECHNOCHROM® C1-Inh. KIT 30T.

REF 5345003  
LOT 0J61B00.01  
⌚ 31-Oct-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	Coagulation Control N	1P37B00	30-Sep-2017
1x1mL	Coagulation Control A	3P38B00	30-Sep-2017
1x0.825mL	Coagulation Reference	1R51B00	31-Oct-2016
1x3mL	C1 Esterase	7B62B00	31-Jan-2018
1x3mL	Substrate f. C1-esterase	8C51B00	31-Mar-2020
1x25mL	C1-Inh Probenpuffer A	9M51B00	30-Jun-2020
1x20mL	C1-Inh Reaktionspuffer B	9N51B00	31-Jul-2020

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B00	77.3 – 128.8%	106.6	OK
	CCA LOT 3P38B00	29.3 – 48.8%	35.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Christoph ENNE

03-Mar-2016  
Datum / Date:

Freigabe von  
Approved by

Lucas ROSENMAYER

03. MRZ. 2016  
Datum / Date: