
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7E00AZ.01
---	---	-------------

PROTAC (Protein C Activator) 5x1mL

REF 5346212
LOT 7E31B00.01
 31-Jan-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability Ceveron® alpha	CCN (1P14B00) Protein C chromogen	77.0 – 128.0% 119.2% OK
	CCA (3P24B00) Protein C chromogen	14.0 – 23.3% 22.0% OK

**Durchführung von
Prepared by**

Alessandro BENIGNI


03-JUL-2013
Datum / Date:

**Freigabe von
Approved by**

Harald SEVCIK


03. JULI 2013
Datum / Date: