



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ0Z00AZ.02


TECHNOCHROM[®] FXIII 3 x 3ml Kit
TECHNOCHROM[®] FXIII 3 x 3ml Set

REF 5360010

LOT 0Z41B00.01

Herstelldatum /
Date of manufacturing 11-Mar-2014

 28-Feb-2015

Inhalt	Reagenz	LOT	
3 x 3 ml	FXIII Activator Reagent	0389100/1	28-Feb-2015
3 x 3 ml	FXIII Detection Reagent	0389100/2	28-Feb-2015
3 x 3 ml	FXIII NADH	0389100/3	28-Feb-2015
3 x 1 ml	FXIII Inhibitor Reagent	0389100/4	28-Feb-2015
1 x 6 ml	Stabilizer solution	0389100/5	28-Feb-2015

Test		Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1R31B00	65% – 109%	90.70%
	Coagulation Control A Lot: 3P24B00	29.3% – 48.8%	38.37%

Durchführung von
Prepared by

Michael RIHA 

12-Mär-2014
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

12-Mär-2014
Datum / Date: