



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ0Z00AZ.02


TECHNOCHROM[®] FXIII 3 x 3ml Kit
TECHNOCHROM[®] FXIII 3 x 3ml Set

REF 5360010

LOT 0Z43B00.01

Herstelldatum /
Date of manufacturing 21-Jul-2014

 31-Jul-2015

Inhalt	Reagenz	LOT	
3 x 3 ml	FXIII Activator Reagent	0656100/1	31-Jul-2015
3 x 3 ml	FXIII Detection Reagent	0656100/2	31-Jul-2015
3 x 3 ml	FXIII NADPH	0656100/3	31-Jul-2015
3 x 1 ml	FXIII Inhibitor Reagent	0656100/4	31-Jul-2015
1 x 6 ml	Stabilizer solution	0656100/5	31-Jul-2015

Test		Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P44B00	63.8% – 106.3%	80.8%
	Coagulation Control A Lot: 3P0AB00	28% – 46%	31.4%

Durchführung von
Prepared by

Michael RIHA 

22-Jul-2014
Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI 

22-Jul-2014
Datum / Date: