

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0Z00AZ.02
---	---	-------------


## TECHNOCHROM® FXIII 3 x 3mL


**REF**            5360010  
**LOT**            0Z51B00.01  
            31-May-2016

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
3 x 3 mL	FXIII Activator Reagent	KTM1224100	31-May-2016
3 x 3 mL	FXIII Detection Reagent	KTM1224100	31-May-2016
3 x 3 mL	FXIII NADPH	KTM1224100	31-May-2016
3 x 1 mL	FXIII Inhibitor Reagent	KTM1224100	31-May-2016
1 x 6 mL	Stabilizer solution	KTM1224100	31-May-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Coagulation Control N (1P37B00)	60.8 – 101.3 %	82.76	OK
	Coagulation Control A (3P38B00)	28.5 – 47.5 %	45.33	OK


Durchführung von  
Prepared by

  
\_\_\_\_\_

  
Datum / Date: \_\_\_\_\_

Freigabe von  
Approved by

 **Fikret NASUFI**  
\_\_\_\_\_

  
Datum / Date: \_\_\_\_\_