



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0Z00AZ.02

TECHNOCHROM® FXIII 3 x 3mL

REF 5360010  
LOT 0Z52B00.01  
⌚ 31-Aug-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
3 x 3 ml	FXIII Activator Reagent	KTM1355100/1	31-Aug-2016
3 x 3 ml	FXIII Detection Reagent	KTM1355100/2	31-Aug-2016
3 x 3 ml	FXIII NADPH	KTM1355100/3	31-Aug-2016
3 x 1 ml	FXIII Inhibitor Reagent	KTM1355100/4	31-Aug-2016
1 x 6 ml	Stabilizer solution	KTM1355100/5	31-Aug-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Coagulation Control N (1P37B00)	60.8 – 101.3 %	82.3	OK
	Coagulation Control A (3P38B00)	28.5 – 47.5 %	39.6	OK

Durchführung von  
Prepared by



11-Sep-2015  
Datum / Date:

Freigabe von  
Approved by

**Fikret NASUFI**  


11-Sep-2015  
Datum / Date: