

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00AZ.02
---	---	-------------

TECHNOCHROM[®] FXIII 3 x 3mL

REF 5360010
LOT 0Z53B00.01
 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
3 x 3 mL	FXIII Activator Reagent	KTM1414100	30-Nov-2016
3 x 3 mL	FXIII Detection Reagent	KTM1414100	30-Nov-2016
3 x 3 mL	FXIII NADPH	KTM1414100	30-Nov-2016
3 x 1 mL	FXIII Inhibitor Reagent	KTM1414100	30-Nov-2016
1 x 6 mL	Stabilizer solution	KTM1414100	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Coagulation Control N (1P37B00)	60.8 – 101.3 %	86.45	OK
	Coagulation Control A (3P38B00)	28.5 – 47.5 %	39.09	OK

Durchführung von
Prepared by

Michael KAFKA 

01-DEC-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

01-DEC-2015
Datum / Date: