	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9G0C.01
---	---	-----------

**Citrat-Natriumchlorid-Puffer 25mL**

REF      5400047  
LOT      9G51BA0.01  
      30-Sep-2020


Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
pH	7.1 – 7.6	7.3	OK	
Natriumcitratgehalt Sodiumcitrate content	5.1 – 5.7g/L	5.6	OK	
Verwendbarkeit, FI Ceveron® alpha	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P37B00	73.5 – 122.5%	111.3	OK
Usability, FI Ceveron® alpha	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	12.8 – 21.3%	20.8	OK

**Durchführung von  
Prepared by**

  
Monika WEINBAUER

22-Oct-2015  
Datum / Date:

**Freigabe von  
Approved by**

  
Lavdrim NASUFI

22-Oct-2015  
Datum / Date: