
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9J00AZ.02
---	---	-------------


Imidazol Puffer 25 mL

REF 5410008
LOT 9J34BE0.01
 30-JUN-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.6	7.49	OK
Kontrollwiederfindung Recovery of controls	CCN LOT 1P2AB00	63.0 – 105.0%	82.1	OK
	CCA LOT 3P24B00	11.9 – 19.9%	15.8	OK
Verwendbarkeit Usability Ceveron® alpha	TCN (1P2FB00)	2.88 – 4.80 g/L	3.32 g/L	OK
	TCA (3P31B00)	1.11 – 1.85 g/L	1.50 g/L	OK

**Durchführung von
Prepared by**

Lucas HOŁOVICZ



07-AUG-2013
Datum / Date:

**Freigabe von
Approved by**



07. AUG. 2013
Datum / Date: