

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9J00AZ.01
---	---	-------------

Imidazol Puffer 50 mL

REF 5410010
LOT 9J52BD0.01
 30-Apr-2020

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result			
pH	7.2 – 7.6	7.5	OK		
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen , Clauss method	CCN LOT 1P37B00	2.52 – 3.78 g/L	3.64	OK
		CCA LOT 3P38B00	0.93 – 1.39 g/L	1.22	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

Durchführung von
Prepared by

Lavdrim NASUFI


09-Jun-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI


09-Jun-2015
Datum / Date: