



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ9J00AZ.01

Imidazol Puffer 90 mL

REF 5410012  
LOT 9J36BC0.02  
⌚ 30-Sep-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
pH	7.2 – 7.6	7.5	OK	
Verwendbarkeit, Fibrinogen, Clauss method	TCN LOT 1A31B02	2.70 – 4.00g/L	3.46	OK
Usability, Fibrinogen, Clauss method Ceveron <sup>®</sup> alpha	CCA LOT 3P38B00	0.93 – 1.39g/L	1.04	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von  
Prepared by

Alessandro BENIGNI

20-Mar-2015  
Datum / Date:

Freigabe von  
Approved by

Lavdrim NASUFI

20-Mar-2015  
Datum / Date: