
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9J00AZ.02
---	---	-------------

Imidazol Puffer 500 mL

REF 5410014
LOT 9J42BD0.01
 31-Jan-2019


Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.2 – 7.6	7.4	OK
Verwendbarkeit Usability Ceveron [®] alpha Fibrinogen, Clauss Method	TCN LOT 1A31B01	2.7 – 4.0 g/L	3.52 g/L OK
	TCA LOT 3P31B01	1.11 – 1.85 g/L	1.39 g/L OK
Etiketten Überprüfung Label Verification		OK	

Durchführung von
Prepared by

Alessandro BENIGNI


18-Feb-2014
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI


18-Feb-2014
Datum / Date: