



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ9J00AZ.02

Imidazol Puffer 500 mL

REF 5410014
LOT 9J4ABG0.01
🕒 30-Nov-2019

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.2 – 7.6	7.5	OK
Verwendbarkeit Usability Ceveron® alpha Fibrinogen, Clauss Method	CCN LOT 1P44B00	2.96 – 4.44 g/L	3.74 g/L OK
	CCA LOT 3P38B00	0.93 – 1.39 g/L	1.25 g/L OK
Etiketten Überprüfung Label Verification		OK	

Durchführung von
Prepared by

Alessandro BENIGNI

03-Feb-2015
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI

03-Feb-2015
Datum / Date: