



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRA00AZ

TECHNOZYM vWF:Ag ELISA

REF 5450201
LOT RA42B00.01
 31 – Aug – 2015

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	vWF:Ag Plate	RNF41B0	29-Feb-2016
1 x 0.3mL	vWF:Ag Conjugate	RNH41B0	31-Jan-2016
1x 80mL	Wash-Buffer	RNA35B0	31-Dec-2015
1 x 90mL	Incubation Buffer	RNB41B0	13-Jan-2016
1 x 12mL	TMB substrate	RKN33B0	30-Sep-2015
1 x 12mL	Stop solution	RKK33B0	31-Oct-2015
1 x 0,5mL	vWF:Ag Calibrator 1	1ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Calibrator 2	2ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Calibrator 3	3ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Calibrator 4	4ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Calibrator 5	5ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag High Control	AND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Low Control	BND31B0	31-Aug-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
OD	Calibrator 1	> 1.0 1.27 OK
	Calibrator 5	< 0.1 0.003 OK
Verwendbarkeit Usability	High Control (AND31B0)	0.44 – 0.73 U/mL 0.67 OK
	Low Control (BND31B0)	0.16 – 0.36 U/mL 0.27 OK

Durchführung von
Prepared by

Fikret NASUFI
Fikret Nasufi

14. Mär - 2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA
Michael Riha

14. Mär - 2014
Datum / Date: