





# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRA00AZ

## TECHNOZYM vWF:Ag ELISA

**REF** 5450201  
**LOT** RA43B00.01  
 31 – Aug – 2015

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 12	vWF:Ag Plate	RNF42B0	31-Mar-2016
1 x 0.3mL	vWF:Ag Conjugate	RNH41B0	31-Jan-2016
1 x 80mL	Wash-Buffer	RNA41B0	29-Feb-2016
1 x 90mL	Incubation Buffer	RNB42B0	30-Apr-2016
1 x 12mL	TMB substrate	RKN34B0	30-Nov-2015
1 x 12mL	Stop solution	RKK41B0	31-Mar-2016
1 x 0,5mL	vWF:Ag Calibrator 1	1ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Calibrator 2	2ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Calibrator 3	3ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Calibrator 4	4ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Calibrator 5	5ND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag High Control	AND31B0	31-Aug-2015
1 x 0.5mL	vWF:Ag Low Control	BND31B0	31-Aug-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.27	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND31B0)	0.44 – 0.73 U/mL	0.56	OK
	Low Control (BND31B0)	0.16 – 0.36 U/mL	0.26	OK

Durchführung von  
Prepared by

Michael RIHA 

19-Aug-2015  
Datum / Date:

Freigabe von  
Approved by

Alessandro BENIGNI 

19-Aug-2015  
Datum / Date: