



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRB00AZ

TECHNOZYM vWF:Ag ELISA Kit 96T

REF 5450201

LOT RA44B00.01

⌚ 31-Jan-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:Ag Plate	RNF43B0	31-Oct-2016
1 x 0.3mL	Ag Conjugate	RNH41B0	31-Jan-2016
1x 80mL	Wash buffer Concentrate	RNA42B0	31-Aug-2016
1 x 90mL	Incubation buffer	RNB43B0	30-Sep-2016
1 x 12mL	TMB Substrate	RKN41B0	31-Aug-2016
1 x 12mL	Stop solution	RKK42B0	30-Sep-2016
1 x 0.5mL	vWF Calibrator 1	1ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 2	2ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 3	3ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 4	4ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 5	5ND31B1	30-Nov-2016
1 x 0.5mL	vWF High Control	AND31B1	30-Nov-2016
1 x 0.5mL	vWF Low Control	BND31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.12	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND31B1)	0.44 – 0.73 IU/mL	0.59	OK
	Low Control (BND31B1)	0.16 – 0.36 IU/mL	0.25	OK

Durchführung von
Prepared by

Michael RIHA

20-Jan-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

20-Jan-2015
Datum / Date: