



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRB00AZ

TECHNOZYM vWF:CBA ELISA Kit 96T

REF 5450301
LOT RB3AB00.01
 31 - Aug - 2015

Inhalt Content	Reagenz Reagent	LOT	
1 x	CBA Plate	RNC37B0	30-Nov-2015
1 x 0.3mL	CBA Conjugate	RNE32B0	31-Aug-2015
1x 80mL	Wash buffer Concentrate	RNA34B0	30-Sep-2015
1 x 90mL	Incubation buffer	RNB33B0	30-Sep-2015
1 x 12mL	TMB Substrat	RKN33B0	30-Sep-2015
1 x 12mL	Stop solution	RKK33B0	31-Oct-2015
1 x 0.5mL	Calibrator 1 vWF	1ND31B0	31-Aug-2015
1 x 0.5mL	Calibrator 2 vWF	2ND31B0	31-Aug-2015
1 x 0.5mL	Calibrator 3 vWF	3ND31B0	31-Aug-2015
1 x 0.5mL	Calibrator 4 vWF	4ND31B0	31-Aug-2015
1 x 0.5mL	Calibrator 5 vWF	5ND31B0	31-Aug-2015
1 x 0.5mL	Control High vWF	AND31B0	31-Aug-2015
1 x 0.5mL	Control Low vWF	BND31B0	31-Aug-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
OD	Calibrator 1	> 1.0 1.352 OK
	Calibrator 5	< 0.1 0.004 OK
Verwendbarkeit Usability	High Control (AND31B0)	0.48 – 0.80 U/mL 0.671 OK
	Low Control (BND31B0)	0.16 – 0.38 U/mL 0.289 OK

Durchführung von
Prepared by

Michael RIHA

16. Dec. 2013
Datum / Date:

Freigabe von
Approved by

16. Dec. 2013
Datum / Date: