



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRB00AZ

TECHNOZYM vWF:CBA ELISA Kit 96T

REF 5450301
 LOT RB41B00.01
 ⌚ 31 - Aug - 2015

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	CBA Plate	RNC38B0	31-Dec-2015
1 x 0.3mL	CBA Conjugate	RNE32B0	31-Aug-2015
1x 80mL	Wash buffer Concentrate	RNA35B0	31-Dec-2015
1 x 90mL	Incubation buffer	RNB33B0	30-Sep-2015
1 x 12mL	TMB Substrat	RKN33B0	30-Sep-2015
1 x 12mL	Stop solution	RKK33B0	31-Oct-2015
1 x 0.5mL	Calibrator 1 vWF	1ND31B0	31-Aug-2015
1 x 0.5mL	Calibrator 2 vWF	2ND31B0	31-Aug-2015
1 x 0.5mL	Calibrator 3 vWF	3ND31B0	31-Aug-2015
1 x 0.5mL	Calibrator 4 vWF	4ND31B0	31-Aug-2015
1 x 0.5mL	Calibrator 5 vWF	5ND31B0	31-Aug-2015
1 x 0.5mL	Control High vWF	AND31B0	31-Aug-2015
1 x 0.5mL	Control Low vWF	BND31B0	31-Aug-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.2	OK
	Calibrator 5	< 0.1	0.03	OK
Verwendbarkeit Usability	High Control (AND31B0)	0.48 – 0.80 U/mL	0.70	OK
	Low Control (BND31B0)	0.16 – 0.38 U/mL	0.25	OK

Durchführung von
Prepared by

Fikret NASUFI

Fikret Nasufi

31-JAN-2014

Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER

31-JAN-2014

Datum / Date: