



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRB00AZ

TECHNOZYM vWF:CBA Collagen Typ III ELISA Kit 96T

REF 5450301

LOT RB49B00.01

⌚ 31-Aug-2015

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:CBA Plate Collagen Typ III	RNC45B0	30-Sep-2016
1 x 0.3mL	CBA Conjugate	RNE41B0	31-May-2016
1x 80mL	Wash buffer Concentrate	RNA42B0	31-Aug-2016
1 x 90mL	Incubation buffer	RNB42B0	30-Apr-2016
1 x 12mL	TMB Substrate	RKN34B0	30-Nov-2015
1 x 12mL	Stop solution	RKK42B0	30-Sep-2016
1 x 0.5mL	vWF Calibrator 1	1ND31B0	31-Aug-2015
1 x 0.5mL	vWF Calibrator 2	2ND31B0	31-Aug-2015
1 x 0.5mL	vWF Calibrator 3	3ND31B0	31-Aug-2015
1 x 0.5mL	vWF Calibrator 4	4ND31B0	31-Aug-2015
1 x 0.5mL	vWF Calibrator 5	5ND31B0	31-Aug-2015
1 x 0.5mL	vWF High Control	AND31B0	31-Aug-2015
1 x 0.5mL	vWF Low Control	BND31B0	31-Aug-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.43	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND31B0)	0.48-0.80 IU/mL	0.80	OK
	Low Control (BND31B0)	0.16-0.38 IU/mL	0.34	OK

Durchführung von
Prepared by

Fikret NASUFI

08-Oct-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA

08-OCT-2014
Datum / Date: