



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRB00AZ

TECHNOZYM® vWF:CBA Collagen Typ III ELISA Kit 96T

REF 5450301

LOT RB51B00.02

⌚ 31-May-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:CBA Plate Collagen Typ III	RNC46B0	31-Dec-2016
1 x 0.3mL	CBA Conjugate	RNE41B0	31-May-2016
1x 80mL	Wash buffer Concentrate	RNA43B0	31-Dec-2016
1 x 90mL	Incubation buffer	RNB43B0	30-Sep-2016
1 x 12mL	TMB Substrate	RKN41B0	31-Aug-2016
1 x 12mL	Stop solution	RKK43B0	31-Dec-2016
1 x 0.5mL	vWF Calibrator 1	1ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 2	2ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 3	3ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 4	4ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 5	5ND31B1	30-Nov-2016
1 x 0.5mL	vWF High Control	AND31B1	30-Nov-2016
1 x 0.5mL	vWF Low Control	BND31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.43	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND31B1)	0.48 – 0.80 IU/mL	0.65	OK
	Low Control (BND31B1)	0.16 – 0.38 IU/mL	0.26	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

25-Feb-2015  
Datum / Date:

Freigabe von  
Approved by

Michael RIHA *Michael Riha*

25-Feb-2015  
Datum / Date: