



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRB00AZ

TECHNOZYM vWF:CBA ELISA Kit 96T

REF 5450301

LOT RB58B00.01

30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x	vWF:CBA Plate Collagen Typ III	RNC53B0	30-Jun-2017
1 x 0.3mL	CBA Conjugate	RNE52B0	30-Jun-2017
1x 80mL	Wash buffer Concentrate	RNA51B0	30-Apr-2017
1 x 90mL	Incubation buffer	RNB51B0	28-Feb-2017
1 x 12mL	TMB Substrate	RKN51B0	30-Apr-2017
1 x 12mL	Stop solution	RKK43B0	31-Dec-2016
1 x 0.5mL	vWF Calibrator 1	1ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 2	2ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 3	3ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 4	4ND31B1	30-Nov-2016
1 x 0.5mL	vWF Calibrator 5	5ND31B1	30-Nov-2016
1 x 0.5mL	vWF High Control	AND31B1	30-Nov-2016
1 x 0.5mL	vWF Low Control	BND31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.50	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND31B1)	0.48 – 0.80 IU/mL	0.68	OK
	Low Control (BND31B1)	0.16 – 0.38 IU/mL	0.27	OK

Durchführung von
Prepared by

Fikret NASUFI

16-Jul-2015
Datum / Date:

Freigabe von
Approved by

[Signature]

16-Jul-2015
Datum / Date: