



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRK00AZ

TECHNOZYM vWF:CBA Collagen Typ I ELISA Kit 96T

REF 5450311

LOT RK42B00.02

30-Aug-2015

Inhalt Content	Reagenz Reagent	LOT	
1 x	vWF:CBA Plate Collagen Typ VI	RNM41B0	31-Aug-2016
1 x 0.3mL	CBA Conjugate	RNE41B0	31-Mai-2016
1x 80mL	Wash buffer Concentrate	RNA42B0	31-Aug-2016
1 x 90mL	Incubation buffer	RNB42B0	30-Apr-2016
1 x 12mL	TMB Substrate	RKN34B0	30-Nov-2015
1 x 12mL	Stop solution	RKK41B0	31-Mar-2016
1 x 0.5mL	vWF:CBA Calibrator 1	1ND31B0	31-Aug-2015
1 x 0.5mL	vWF:CBA Calibrator 2	2ND31B0	31-Aug-2015
1 x 0.5mL	vWF:CBA Calibrator 3	3ND31B0	31-Aug-2015
1 x 0.5mL	vWF:CBA Calibrator 4	4ND31B0	31-Aug-2015
1 x 0.5mL	vWF:CBA Calibrator 5	5ND31B0	31-Aug-2015
1 x 0.5mL	vWF:CBA High Control	AND31B0	31-Aug-2015
1 x 0.5mL	vWF:CBA Low Control	BND31B0	31-Aug-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.24	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND31B0)	0.48 – 0.80 IU/mL	0.59	OK
	Low Control (BND31B0)	0.16 – 0.36 IU/mL	0.23	OK

Durchführung von
Prepared by

Michael RIHA

03-Oct-2016
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

03-Oct 2016
Datum / Date: