



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRK00AZ

TECHNOZYM[®] vWF:CBA Collagen Typ I ELISA Kit 96T

REF 5450311

LOT RK52B00.01

30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x	vWF:CBA Plate Collagen Typ I	RNM51B0	31-Aug-2017
1 x 0.3mL	vWF:Ag Conjugate	RNH51B0	28-Feb-2017
1x 80mL	Wash buffer Concentrate	RNA51B0	30-Apr-2017
1 x 90mL	Incubation buffer	RNB51B0	28-Feb-2017
1 x 12mL	TMB Substrate	RKN51B0	30-Apr-2017
1 x 12mL	Stop solution	RKK51B0	30-Jun-2017
1 x 0.5mL	vWF:CBA Calibrator 1	1ND31B1	30-Nov-2016
1 x 0.5mL	vWF:CBA Calibrator 2	2ND31B1	30-Nov-2016
1 x 0.5mL	vWF:CBA Calibrator 3	3ND31B1	30-Nov-2016
1 x 0.5mL	vWF:CBA Calibrator 4	4ND31B1	30-Nov-2016
1 x 0.5mL	vWF:CBA Calibrator 5	5ND31B1	30-Nov-2016
1 x 0.5mL	vWF:CBA High Control	AND31B1	30-Nov-2016
1 x 0.5mL	vWF:CBA Low Control	BND31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.55	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (AND31B1)	0.48 – 0.80 IU/mL	0.69	OK
	Low Control (BND31B1)	0.16 – 0.38 IU/mL	0.26	OK

Durchführung von
Prepared by

15-SEP-2016
Datum / Date:

Freigabe von
Approved by

15-Sep-2016
Datum / Date: