





**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRP00AZ

TECHNOZYM ADAMTS-13 INH

REF 5450401
LOT RP44B00.01
 30 - JUN - 2015

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 INH Plate	RPS43B0	31-Mar-2016
1 x 0.3mL	Adamts13 INH Conjugate	RPV41B0	31-JAN-2016
1x 80mL	Wash buffer Concentrate	RNA35B0	31-DEC-2015
1 x 90mL	Incubation buffer	RNB41B0	31-JAN-2016
1 x 12mL	TMB Substrate	RKN33B0	30-SEP-2015
1 x 12mL	Stop solution	RKK33B0	31-OCT-2015
1 x 0,5mL	Adamts13 INH Calibrator 1	1PT31B0	30-JUN-2015
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT31B0	30-JUN-2015
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT32B0	31-JUL-2015
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT31B0	31-JUL-2015
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT31B0	31-JUL-2015
1 x 0.5mL	Adamts13 INH High Control	RPY31B0	31-JUL-2015
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B0	31-JUL-2015


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.648	OK
	Calibrator 5	< 0.1	0.012	OK
Verwendbarkeit Usability	High Control (RPY31B0)	53.1 – 79.6	66.231	OK
	Low Control (RPZ31B0)	4.21 – 9.83	8.400	OK

Durchführung von
Prepared by

Fikret NASUFI


24-May-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA


24-MAY-2014
Datum / Date: