



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM ADAMTS-13 INH ELISA Kit 96T

REF 5450401
LOT RP4FB00.01
⌚ 31-Aug-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 6	Adamts13 INH Plate	RPS4BB0	30-Nov-2016
1 x 0.3mL	Adamts13 INH Conjugate	RPV43B0	31-Oct-2016
1x 80mL	Wash buffer Concentrate	RNA42B0	31-Aug-2016
1 x 90mL	Incubation buffer	RNB43B0	30-Sep-2016
1 x 12mL	TMB Substrate	RKN41B0	31-Aug-2016
1 x 12mL	Stop solution	RKK42B0	30-Sep-2016
1 x 0,5mL	Adamts13 INH Calibrator 1	1PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT32B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH High Control	RPY31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
OD	Calibrator 1	> 1.0
	Calibrator 5	< 0.1
Verwendbarkeit Usability	High Control (RPY31B1)	53.10 – 79.64 U/mL
	Low Control (RPZ31B1)	4.21 – 9.83 U/mL

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

13-Jan-2015
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ *Lucas HoloVICZ*

13-Jan-2015
Datum / Date: