



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T

REF 5450401

LOT RP55B00.01

🕒 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x 12	Adamts13 INH Plate	RPS52B0	31-Dec-2017
1 x 0.3mL	Adamts13 INH Conjugate	RPV51B0	31-May-2017
1x 80mL	Wash buffer Concentrate	RNA52B0	31-Oct-2017
1 x 90mL	Incubation buffer	RNB52B0	31-Jul-2017
1 x 12mL	TMB Substrate	RKN51B0	30-Apr-2017
1 x 12mL	Stop solution	RKK51B0	30-Jun-2017
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT32B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH High Control	RPY31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.48	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY31B1)	53.10 – 79.64 U/mL	58.24	OK
	Low Control (RPZ31B1)	4.21 – 9.83 U/mL	5.76	OK

Durchführung von
Prepared by

Michael KAFKA

10-DEC-2017
Datum / Date:

Freigabe von
Approved by

Fikret Nasufi
Fikret NASUFI

10-Dec-2015
Datum / Date: